

TECHNICAL REPORT

TITLE : **Technical Report on Spectroscopy Amplifier**

AUTHORS : **S.Venkataramanan, Arti Gupta**

CATEGORY : **Instrumentation**

REFERENCE NO : **NSC/TR/SV/2002-03/29**

NUCLEAR SCIENCE CENTRE

(An Autonomous Inter-University Centre of UGC)

Post Box No.: 10502, Aruna Asaf Ali Marg,

New Delhi 110067 (India) Phone: 2689 3955, 2689 2603, 2689 2601

Fax: 091-11-2689 3666 Email: info@nsc.ernet.in

TECHNICAL REPORT

ON SPECTROSCOPY AMPLIFIER

CONTENTS

- 1. Abstract**
- 2. Acknowledgment**
- 3. Specifications**
- 4. Introduction**
- 5. Principle of operations**
- 6. Pile Up rejection**
- 7. Performances**
- 8. Assembly Procedure**
- 9. References**
- 10. Schematic, Block and signal level diagrams**
- 11. Bill of material**
- 12. PCB artwork TOP, BOTTOM layers**
- 13. Silk screen TOP, BOTTOM layers**
- 14. Drill drawing**

Prepared by

**ELECTRONICS LABORATORY
NUCLEAR SCIENCE CENTRE,
P.B.10502,
NEW DELHI 110067.**

Technical Report on Spectroscopy Amplifier

S.Venkataramanan*, Arti Gupta
Electronics Laboratory, Nuclear Science Centre,
P.B.10502, Aruna Asaf Ali Marg,
New Delhi 110067.
*email: venkat@nsc.ernet.in

Abstract

As a part of on going nuclear electronics development activities for INGA project at NSC, we have successfully developed a high density Spectroscopy amplifier for Clover detector with SMD technology. It essentially consists of active Pole-Zero (P/Z) compensation, gain stages, active shaping, Gated BLR correction, Pileup Rejection (PUR) logic and buffer circuits. It has the options for remote P/Z adjustment and threshold setting for base line restorer (BLR) correction.

Various prototypes have been tested with Clover detectors. In this report the principle of operation of various circuit blocks are explained along with representative test point signals. Assembly and troubleshooting procedures are given. The amplifier specifications and test results are given and possible improvements are also suggested.

Acknowledgment

We would like to thank engineers from GIP, Ganil, France for their constant support in simulating various circuit blocks and fruitful discussions. We also would like to thank INGA project group at NSC for specifying and evaluating spectroscopy amplifiers and providing funds for development. At last, our sincere thanks to Prof. G.K.Mehta, Dr.Amit Roy and Ajith Kumar.B.P, for their constant encouragement and providing the necessary infrastructure inorder to complete this project successfully.

Introduction

A nuclear spectroscopy (shaping) amplifier is a signal amplifier with complete DC correction control loop, while shaping the nuclear detector pulses of short rise time and long fall time into a finite width semi-Gaussian shape, which can be digitized with ADC. A typical modern shaping amplifier consists of a P/Z correction, gain stages, active integration, Active gated base line restorer and coaxial cable driver while providing convenient synchronous (BUSY) and pile up reject (PUR) logic pulses.

We have incorporated all above features in our shaping amplifier, in which the shaping time constant is fixed. The Pole-Zero adjustment and manual threshold settings can be done from either front panel or thorough remotely controlled digital to analog converter(DAC). Three gain selections are possible with onboard jumper selection. In order to reduce dead time related to overload, we have incorporated limiters in gain stages for fast recovery. The Quasigaussian pulse shaping is achieved with early passive differentiation stage and 2 stages of second order active integration. Slow base line shift due to power line pickup or offset errors due to operational amplifier are corrected with sophisticated gated base line restorer. The piled up events, which would cause poor energy resolution can be rejected with PUR logic pulse at ADC level. A 50 ohm coaxial buffer amplifier is wired in a composite amplifier configuration in order to achieve good high frequency and DC responses.

The entire amplifier is wired as a daughter card on a double sided glass epoxy PCB with surface mount components to achieve required high density.

Principle of Operation

The shaping amplifier schematic is divided into Analog and digital sections and a block diagram is also attached for quick reference. The nuclear radiation pulses received from HPGe which have nominal amplitude of -200mV/MeV and decay time of $50\mu\text{S}(\pm 5\%)$. The early differentiation is done with passive RC components for $3\mu\text{S}$ (C3,C4,R13). The Coarse Pole-Zero(P/Z) correction is done with (R19, C3,C4) and fine adjustment with in $\pm 5\%$ can be done from front panel potentiometer (R14) when remote control is not desired. The remote P/Z adjustment is implemented with wide bandwidth Analog multiplier, which replaces potentiometer. It can also accept control voltage (P/Z_adj.) from remote DAC. The multiplier output can also be used for P/Z compensation in timing channel (TFA).

The differentiated pulse is amplified with low noise high quality operational amplifier (U2). The feedback resistor is varied (jumper selected) for different gain settings (2MeV, 4MeV, 6MeV) as per the user requirement. The pulse is further amplified with low noise operational amplifier (U1). These two stages are wired with clamping circuit in feedback path to speedup overload recovery. The amplified pulses are shaped with 2 stages of 2nd order active integration (U3 A&B), where RC components are chosen for required shaping time constant $3\mu\text{S}$, to provide smooth quasis Gaussian shaped pulses. While amplifying the signals in previous stages, any offset or DC shift voltages generated are to be corrected for obtaining correct energy information. This is achieved with sophisticated gated base line restorer (U6A&B, U7, C19, Q1) which is essentially a closed control loop, with active integration in its feedback path to provide overall differentiation action.

Since the BLR correction cannot be done during presence of pulses, the integrator is gated with elaborate gating circuit (U8 to U11) and level converter. The gate for BLR is generated from Look ahead command, which indicates the presence of energy signal and Wrap around command, which indicates the end of energy pulse.

The look ahead command is generated in U9A, which is an ultra fast comparator (TTL) which compares, the differentiated (300nS) BLR corrected input pulse (BG_IN) and external BLR threshold (U8), which is usually set above input noise level. The presence of energy pulse is indicated by activating a monoshot (U10). The Wrap around command is generated by comparing (U9B) BLR threshold and V_SHAPE, the shaped output pulse, to indicate the end of energy pulse, where the BLR threshold is set just above output noise level. These two commands are processed (U11) to generate a BLR_CONT control command, which would switch the transconductance amplifier (U7) accordingly in BLR control loop. The "BUSY" logic pulse essentially indicates the presence of energy pulse, which is often used to synchronize the ADC.

Pile-Up Rejection

In a HPGe detector, when two gamma rays arrive within the width of the shaping amplifier output pulse, the respective amplifier pulses pileup and generate a distorted pulse and cause poor energy resolution. In such cases, where charge collection time is shorter than peaking time of the amplifier, a PUR pulse can be used to reject such pulses from being processed further.

In this amplifier, the presence of energy signal is indicated by Look ahead command, CLK, in turn a 20 μ S width pulse is generated which is inspection interval for PUR. The associated D flip-flop, would output pulse having width of inspection interval as PUR logic pulse, upon receiving another Look ahead command within the inspection interval.

Assembly Procedure

The currently (PT-5_CORR) available PCB is of glass epoxy, double sided with 0.7mm drill PTH having dimension of 4" x 1.75" with all above features. This has additional provision of PUR logic output. All developed prototypes have common PCB foot prints in order to use them conveniently.

It is recommended to have solder mask and silk screen printed on both sides for easy assembly as well to protect it from solder bridges etc.,. Use of 0.8mm sharp solder tip, IC solder tips are recommended in order to solder narrowly spaced SMT devices. SMT devices shall be picked only by fine quality tweezers. While soldering a magnifier x5 (large) and x12 (eye piece) is used to assure the soldering. It is essential to use solder cleaning liquid with cotton swab to remove dust attracting solder paste.

The PCB shall be checked with magnifiers and multimeter for any unwanted connections and PTHs. Then components shall be soldered in a orderly manner, to start with all low profile chip resistors and capacitors. It is essential to check the impedance between various nodes after soldering resistors, capacitors and inductors. Active components like diodes, transistors and ICS are soldered thereafter. At last tantalum capacitors, connectors, jumpers and any non-SMT devices. All PCBs shall be marked distinctly with unique number for any future references.

References

The following references were proven to be very useful to design and test our shaper.

1. Radiation detection systems, Lecture notes on Pulse Processing techniques in radiation & spectrometry, *S.K.Kataria*, BARC.
2. EG&G ORTEC catalogue
3. Opamps for Everyone, *Ron Moncini*, M/s. TI, USA (SL0D006A)
4. Design with OPAMPS and Analog Integrated Circuits, *Sergio Franco*, McGraw-Hill
5. Techniques for Nuclear & Particle physics Experiments, *W.R.Leo*, Springer Verlag
6. Radiation Detectors Physical Principles & Applications, *Delaney, Finch*, Clarendon Press
7. Spectroscopy Amplifier, Instruction manual, *M/s. SILENA*, Italy.
8. Signal processing for semi-conductor detectors, *F.S.Goulding, A.Landis*, IEEE Transact. NS-29, Vol:3, June 1982.

SPECIFICATIONS: SPECTROSCOPY AMPLIFIER

Input Impedance	~1000 ohm
Pole/Zero adj.	Input pulse having decay time about 50 μ Sec ($\pm 5\%$) can be corrected through potentiometer (FP) or remote controlled through DAC. Zin: 1Kohm. Control voltage not to exceed ± 1 Volt.
Shaping time & type	~3 μ Sec. Fixed, Active integration (4 th order) Quasi Gaussian having peaking time of ~2.4 τ s.
Input signal*	-200mV/MeV is expected to generate +10Volts at the output for 3 gain settings. Not to exceed ± 5 V.
Gain	2MeV, 4MeV and 6 MeV for +10V output. Jumper selection on board. Default is 2 MeV.
BLR threshold	Manual baseline restorer threshold is set through Potentiometer (FP) or remotely. Range is 0 to 500mV when provided. Zin: 1k. Control voltage not to exceed ± 5 V.
Output	Unipolar, Gated BLR DC restored. Width : ~20 μ Sec. Zo: 50 ohms.
BUSY	A TTL negative logic pulse for the duration of presence of output pulse exceeds BLR threshold. Zo: 10 ohms
PUR	Pile up reject signal is a TTL logic signal, with pileup inspection interval of 20 μ Sec. Zo: 10 ohms
Power required	+/-6V, 40mA/20mA +/-12V 40mA/30mA +24V 5mA
Size & Weight	W x H x L : 1.75"x0.5"x 4", 30 grams.

PERFORMANCES:

The module has been subjected to various tests at NSC with ⁶⁰Co and ¹⁵²Eu sources and in beam, in parallel with commercial modules with 8K MCA. The typical results obtained are :

Resolution: 1.3KeV (122KeV), 2.0 KeV (1408 KeV) of ⁶⁰Co @ 9 Kcps.

Integral non-linearity: ± 100 eV for ¹⁵²Eu spectrum, 2MeV/~8V.

Stability: With ⁶⁰Co, no significant shift observed at 6 kcps in 55 hours.





Fig: Photo of Shaper Daughter Card

Fig: Typical BUSY, P/Z Monitor Signals of a Ge. Detector.



Gain Selection Procedure:

The Shaper is designed to work with one of the three different GAIN settings 2MeV*/4MeV/6MeV as per user requirement. The GAIN can be selected on the SHAPER daughter card by the procedure given here.

1. Open the side panel of Clover Electronics Module
2. Locate the SHAPERS in TOP of Mother Board.
3. Identify JUMPER SOCKETS in extreme right corner facing top. (Ref. Photograph)
4. By plugging ' in any on the jumper will select 4MeV.
5. By plugging ' in both jumpers will select 6MeV.

* Default GAIN selection when jumpers not used.

Pole_Zero Adjustment:

The Shaper is designed to work with Eurisy Measures Clover detectors with preamplifiers having $50\mu\text{s}$ ($\pm 10\%$) decay time constant. The Pole_Zero adjustment can be done with front panel PZ_Adj. Potentiometer while monitoring corresponding front panel PZ_ Mon. on a CRO.

BLR Adjustment:

The Shaper is designed to have stable zero reference at all specified working conditions. This is achieved by Gated BLR operation. The required threshold level above system noise is fed through front panel BLR adj. potentiometer. This is set while monitoring front panel BUSY (TTL) signal on a CRO for a minimum count rate when no radiation sources are used. It is essential to set proper Pole Zero adjustment for proper functioning of BLR.

During above procedures, it is recommended to use BUSY signal to trigger CRO.




Fig: Block diagram of a Shaper daughter Card.

Note:

The actual schematics & artworks, Technology can be obtained by writing to Director of this centre.

The module design and content are patented under Indian Patent Law.